

PCB ANALYTICAL RESULTS

Company:	Trinity Analytical Labs., Inc.	Bid Number:	Trinity
Contact:	Contact Name	Phone:	(316) 328-3222
		Fax:	(316) 328-2033
Address:	115 East Fifth Street; P.O. Box A	Date Received:	05/05/2009
	Mound Valley, KS 67354-0000	Date Reported:	06/01/2009
		No. of PCB	13

Client Sample ID	Trinity Sample ID	Sample Matrix ⁽¹⁾	Aroclor Type ⁽²⁾	Result Total PCB ⁽³⁾	Quant. Limit	Rpt. Units
Sample 1	9050505-01	Hydrocarbon		ND	2.0	mg/kg
Sample 2	9050505-02	Hydrocarbon		ND	2.0	mg/kg
Sample 3	9050505-03	Hydrocarbon	1260	6.4	2.0	mg/kg
Sample 4	9050505-04	Hydrocarbon		ND	2.0	mg/kg
Sample 5	9050505-05	Hydrocarbon		ND	2.0	mg/kg
Sample 6	9050505-06	Hydrocarbon		ND	2.0	mg/kg
Sample 7	9050505-07	Hydrocarbon	1260	32	2.0	mg/kg
Sample 8	9050505-08	Hydrocarbon	1260	5.7	2.0	mg/kg
Sample 9	9050505-09	Wipe	1260	1.0	1.0	ug/Wipe
Sample 10	9050505-10	Hydrocarbon		ND	2.0	mg/kg
Sample 11	9050505-11	Wipe		ND	1.0	ug/Wipe
Sample 12	9050505-12	Hydrocarbon		ND	2.0	mg/kg
Sample 13	9050505-13	Hydrocarbon		ND	2.0	mg/kg

NOTES: (1) **Matrix/Method**
 Hydrocarbon - ASTM-4059 (4)
 Solid - EPA 8082
 Surface Wipes - EPA 8082M

Matrix/Method
 Water - EPA 8082
 Air - NIOSH 5503

(2) Aroclors not listed were non-detectable at or above the quantitation limit.
 (3) ND = All aroclors were non-detectable at or above the quantitation limit.
 (4) This method is approved for use under the "Mega Rule" Disposal of Polychlorinated Biphenyls (PCBs); Final Rule, 40 CFR Parts 750 and 761, FR, June 29, 1998, (Volume 63, Number 124), however it is not considered part of NELAC.



Scott A. Popejoy
 Laboratory Manager
 Trinity Analytical Laboratories, Inc.